

Table 3 Sensitivity, specificity and predictive values of rubella IgM enzyme immunoassay (EIA) compare to rubella IgG avidity assay

Test	Sensitivity	Specificity	Positive predictive value	Negative predictive value
	% (95% CI)	% (95% CI)	% (95% CI)	% (95% CI)
Rubella IgM-EIA	76.9 (69.3–84.5)	99.7 (99.3–100.0)	97.8 (96.6–98.6)	96.2 (95.1–97.7)

CI = confidence interval.